



Chelsea Instruments Ltd  
2/3 Central Avenue  
West Molesey  
Surrey KT8 2QZ

Tel: 0181 941 0044  
Fax: 0181 941 9319  
E-Mail: 100731.2345@compuserve.com  
Internet: <http://www.chelsea.co.uk>

## TEST REPORT

This is to certify that Alphatracka serial number 161/2642/004 has been pressure tested for one hour at a pressure of 60bar, with no apparent leakage.

The design of Alphatracka has previously been verified as suitable for continuous use at a maximum depth of 6000m.

Signed

Date

3<sup>rd</sup> January 1997

S Oldfield  
Calibration Manager

### Directors:

D. J. Chatterton MBE Managing Director  
M. B. Kersten FCA  
J. P. Vessey PhD ARCS  
D. Griffiths PhD ARCS  
R. Burt BSc  
B. A. Wardhaugh MIMgt



Certificate No FM25404

Registered No: 552429  
Registered at above address