

Formula

Type: Methane Sensor METS

Serial number: **G63-E277**

$$c = \exp \left[2,025 * \ln \left\{ \left(0,300 + 0,977 * \exp \frac{-V_t}{0,728} \right) * \left(\frac{1}{0,19 + 1,14 * V_{CH_4}} - \frac{1}{-2,149 + 11,871 * \exp \frac{-V_t}{1,271}} \right) \right\} \right]$$

$$t = (V_t * 22,80) - 2,17$$

O₂ correction:

$$c_{CH_4}(corr) = \left(-0,326 + 0,360 * \exp \frac{c_{O_2}}{77,779} \right) * c_{CH_4}$$

c_{CH_4} = methane concentration [$\mu\text{mol/l}$]

t = gas temperature [$^{\circ}\text{C}$]

V_{CH_4} = methane voltage [V]

V_t = temperature voltage [V]

c_{O_2} = O₂ concentration [%]

Methane range: 10 nmol/l - 1 $\mu\text{mol/l}$

Temperature range: 10 – 30 $^{\circ}\text{C}$

Calibrator : J.G.

Date :