

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5166
 CALIBRATION DATE: 21-Apr-07

SBE16plus TEMPERATURE CALIBRATION DATA
 ITS-90 TEMPERATURE SCALE

ITS-90 COEFFICIENTS

a0 = 1.286185e-003
 a1 = 2.601005e-004
 a2 = 8.796172e-008
 a3 = 1.413433e-007


BATH TEMP (ITS-90)	INSTRUMENT OUTPUT(n)	INST TEMP (ITS-90)	RESIDUAL (ITS-90)
1.0000	602640.600	1.0000	0.0000
4.4999	533582.900	4.4999	-0.0000
15.0000	363280.400	15.0000	-0.0000
18.5000	317851.800	18.5001	0.0001
23.9998	256413.900	23.9998	-0.0000
29.0000	209838.300	29.0000	-0.0000
32.5000	181796.400	32.5001	0.0000

$MV = (n - 524288) / 1.6e+007$

$R = (MV * 2.900e+009 + 1.024e+008) / (2.048e+004 - MV * 2.0e+005)$

Temperature ITS-90 = $1 / \{a_0 + a_1[\ln(R)] + a_2[\ln^2(R)] + a_3[\ln^3(R)]\} - 273.15$ (°C)

Residual = instrument temperature - bath temperature



SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 5166
 CALIBRATION DATE: 21-Apr-07


SBE16plus CONDUCTIVITY CALIBRATION DATA
 PSS 1978: C(35,15,0) = 4.2914 Siemens/meter

COEFFICIENTS:

g = -1.018693e+000	CPcor = -9.5700e-008
h = 1.593015e-001	CTcor = 3.2500e-006
i = -5.731472e-004	
j = 6.936952e-005	


BATH TEMP (ITS-90)	BATH SAL (PSU)	BATH COND (Siemens/m)	INST FREQ (Hz)	INST COND (Siemens/m)	RESIDUAL (Siemens/m)
22.0000	0.0000	0.00000	2536.82	0.0000	0.00000
1.0000	34.8329	2.97724	5026.24	2.9772	-0.00000
4.4999	34.8124	3.28438	5215.43	3.2844	0.00000
15.0000	34.7688	4.26640	5778.11	4.2664	-0.00001
18.5000	34.7592	4.61161	5963.04	4.6116	0.00001
23.9998	34.7485	5.16963	6250.05	5.1696	0.00001
29.0000	34.7423	5.69156	6506.67	5.6915	-0.00002
32.5000	34.7371	6.06374	6683.46	6.0638	0.00001

f = INST FREQ / 1000.0
 Conductivity = (g + hf² + if³ + jf⁴) / (1 + δt + εp) Siemens/meter
 t = temperature[°C]; p = pressure[decibars]; δ = CTcor; ε = CPcor;
 Residual = instrument conductivity - bath conductivity



CALIBRATION COEFFICIENTS		SERIAL NO : 104227	
PRESSURE TRANSDUCER		DATE : 08-31-2006	
MODEL :	PRESSURE RANGE :	TEMP. RANGE :	PORT :
46K-134	0 to 6000 psia	0 to 125 deg C	oil filled

<p>PRESSURE COEFFICIENTS</p> <p>U = temperature (deg C)</p> <p>$C = C_1 + C_2U + C_3U^2$</p> <p>D = $D_1 + D_2U$</p> <p>$T_0 = T_1 + T_2U + T_3U^2 + T_4U^3 + T_5U^4$</p> <p>T = pressure period (μsec)</p> <p>Pressure : (psia)</p> <p>$P = C \left(1 - \frac{T_0^2}{T^2}\right) \left(1 - D \left(1 - \frac{T_0^2}{T^2}\right)\right)$</p>		<table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="text-align: center;">C₁</td><td style="text-align: center;">-29493.83 psia</td></tr> <tr><td style="text-align: center;">C₂</td><td style="text-align: center;">-4.76506E-01 psia/deg C</td></tr> <tr><td style="text-align: center;">C₃</td><td style="text-align: center;">8.19742E-03 psia/deg C²</td></tr> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="text-align: center;">D₁</td><td style="text-align: center;">0.032968</td></tr> <tr><td style="text-align: center;">D₂</td><td style="text-align: center;">0</td></tr> </table> <table border="1" style="width: 100%; border-collapse: collapse;"> <tr><td style="text-align: center;">T₁</td><td style="text-align: center;">30.00898 μsec</td></tr> <tr><td style="text-align: center;">T₂</td><td style="text-align: center;">-5.05845E-04 μsec/deg C</td></tr> <tr><td style="text-align: center;">T₃</td><td style="text-align: center;">3.86287E-06 μsec/deg C²</td></tr> <tr><td style="text-align: center;">T₄</td><td style="text-align: center;">1.94603E-09 μsec/deg C³</td></tr> <tr><td style="text-align: center;">T₅</td><td style="text-align: center;">0</td></tr> </table>	C ₁	-29493.83 psia	C ₂	-4.76506E-01 psia/deg C	C ₃	8.19742E-03 psia/deg C ²	D ₁	0.032968	D ₂	0	T ₁	30.00898 μ sec	T ₂	-5.05845E-04 μ sec/deg C	T ₃	3.86287E-06 μ sec/deg C ²	T ₄	1.94603E-09 μ sec/deg C ³	T ₅	0
C ₁	-29493.83 psia																					
C ₂	-4.76506E-01 psia/deg C																					
C ₃	8.19742E-03 psia/deg C ²																					
D ₁	0.032968																					
D ₂	0																					
T ₁	30.00898 μ sec																					
T ₂	-5.05845E-04 μ sec/deg C																					
T ₃	3.86287E-06 μ sec/deg C ²																					
T ₄	1.94603E-09 μ sec/deg C ³																					
T ₅	0																					
(08-31-2006)																						

<p>PAROSCIENTIFIC, INC. 4500 148th AVENUE N.E. REDMOND, WA. 98052</p>	<p>CUSTOMER : SEABIRD ELECTRONICS, INC.</p> <p style="text-align: right;">  SALES ORDER : 23379 PREPARED BY : T.C. </p>
--	---

CALIBRATION COEFFICIENTSSERIAL NO : **104227**

PRESSURE TRANSDUCER

DATE : **08-31-2006**

MODEL : 46K-134	PRESSURE RANGE : 0 to 6000 psia	TEMP. RANGE : 0 to 125 deg C	PORT : oil filled
---------------------------	---	--	-----------------------------

PRESSURE COEFFICIENTS AT FIXED TEMPERATURE

(only valid at specified temperature)

T = pressure period (μsec)Pressure equation : (psia)

$$P = C \left(1 - \frac{T_0^2}{T^2} \right) \left(1 - D \left(1 - \frac{T_0^2}{T^2} \right) \right)$$

Temperature: **21.0 C**

C (psia)	-29500.22				
D	0.032968				
T ₀ (μsec)	30.00008				

(08-31-2006)

PAROSCIENTIFIC, INC.4500 148th AVENUE N.E.
REDMOND, WA. 98052CUSTOMER : **SEABIRD ELECTRONICS, INC.**SALES ORDER : **23379**PREPARED BY : **T.C.**