



**Sea-Bird Electronics, Inc.**  
1808 136<sup>th</sup> Place NE, Bellevue, Washington 98005 USA  
Website: <http://www.seabird.com>

**FAX: (425) 643-9954**  
Tel: (425) 643-9866  
Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 4/8/02

Description: SBE-37 Microcat

Job Number: 28164

Customer: COMM TECH

### SBE Sensor Information:

Model Number: 37

Serial Number: 2399

### Pressure Sensor Information:

Sensor Type: Druck

Serial Number: 1549

Sensor Rating: 5000

### Pressure Test Protocol:

Low Pressure Test: 50 PSI


Held For 15 Minutes

High Pressure Test: 5000 PSI

Held For 15 Minutes

Passed

Tested By: PCC





**Sea-Bird Electronics, Inc.**  
 1808 136<sup>th</sup> Place NE, Bellevue, Washington 98005 USA  
 Website: <http://www.seabird.com>

**FAX: (425) 643-9954**

Tel: (425) 643-9866  
 Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 4/8/02

Description: SBE-5M Mini-Submersible Pump

Job Number: 28164

Customer: COMM TECH

### SBE Sensor Information:

Model Number: 5M

Serial Number: 0434

### Pressure Sensor Information:

Sensor Type: None

Serial Number: None

Sensor Rating: 0

### Pressure Test Protocol:

Low Pressure Test: 50 PSI

Held For 15 Minutes

High Pressure Test: 10000 PSI

Held For 15 Minutes

Passed

Tested By: ML

High pressure is generally equal to the maximum depth rating of the instrument

