

SEA-BIRD ELECTRONICS, INC.

1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1417
 CALIBRATION DATE: 21-Nov-03

SBE 37 PRESSURE CALIBRATION DATA
 10000 psia S/N 193243

COEFFICIENTS:

PA0 = 3.243866e+000	PTCA0 = -2.774054e+002
PA1 = 4.674189e-001	PTCA1 = 3.598518e-001
PA2 = -5.842712e-008	PTCA2 = -5.528487e-004
	PTCB0 = 2.571000e+001
	PTCB1 = 1.233333e-003
	PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.73	-248.2	19.9	13.64	-0.01
2015.02	4037.5	20.3	2013.76	-0.01
4015.16	8324.9	20.1	4012.67	-0.02
6015.47	12618.7	20.2	6012.32	-0.03
8015.72	16921.0	20.3	8013.74	-0.02
10016.41	21230.6	20.1	10016.58	0.00
10016.41	21230.6	20.1	10016.58	0.00
8015.72	16929.1	20.3	8017.51	0.02
6015.39	12630.4	20.3	6017.72	0.02
4015.13	8336.1	20.2	4017.85	0.03
2014.93	4045.1	20.5	2017.26	0.02
14.62	-244.9	20.5	15.09	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	-240.73	0.00	25.71
29.00	-241.97	30.00	25.75
24.00	-243.54		
18.50	-245.40		
15.00	-246.64		
4.50	-250.20		
1.00	-251.57		

$$y = \text{thermistor output}; t = P\text{TEMPA}0 + P\text{TEMPA}1 * y + P\text{TEMPA}2 * y^2$$


$$x = \text{pressure output} = P\text{TCA}0 + P\text{TCA}1 * t + P\text{TCA}2 * t^2$$

$$n = x * P\text{TCB}0 / (P\text{TCB}0 + P\text{TCB}1 * t + P\text{TCB}2 * t^2)$$

$$\text{pressurc (psia)} = P\text{A}0 + P\text{A}1 * n + P\text{A}2 * n^2$$

Date, Delta P %FS

● 19-Apr-02 0.24
 ▲ 21-Nov-03 0.00



SEA-BIRD ELECTRONICS, INC.
 1808 136th Place N.E., Bellevue, Washington, 98005 USA
 Phone: (425) 643 - 9866 Fax (425) 643 - 9954 Email: seabird@seabird.com

SENSOR SERIAL NUMBER: 1417
 CALIBRATION DATE: 21-Nov-03

SBE 37 PRESSURE CALIBRATION DATA
 10000 psia S/N 193243

COEFFICIENTS:

PA0 = 3.243866e+000
 PA1 = 4.674189e-001
 PA2 = -5.842712e-008

PTCA0 = -2.774054e+002
 PTCA1 = 3.598518e-001
 PTCA2 = -5.528487e-004
 PTCB0 = 2.571000e+001
 PTCB1 = 1.233333e-003
 PTCB2 = 0.000000e+000

PRESSURE SPAN CALIBRATION

PRESSURE PSIA	INST OUTPUT	TEMP ITS90	COMPUTED PRESSURE	ERROR %FS
14.73	-248.2	19.9	13.64	-0.01
2015.02	4037.5	20.3	2013.76	-0.01
4015.16	8324.9	20.1	4012.67	-0.02
6015.47	12618.7	20.2	6012.32	-0.03
8015.72	16921.0	20.3	8013.74	-0.02
10016.41	21230.6	20.1	10016.58	0.00
10016.41	21230.6	20.1	10016.58	0.00
8015.72	16929.1	20.3	8017.51	0.02
6015.39	12630.4	20.3	6017.72	0.02
4015.13	8336.1	20.2	4017.85	0.03
2014.93	4045.1	20.5	2017.26	0.02
14.62	-244.9	20.5	15.09	0.00

THERMAL CORRECTION

TEMP ITS90	INST OUTPUT	TEMP ITS90	SPAN MV
32.50	-240.73	0.00	25.71
29.00	-241.97	30.00	25.75
24.00	-243.54		
18.50	-245.40		
15.00	-246.64		
4.50	-250.20		
1.00	-251.57		

$$y = \text{thermistor output}; t = P\text{TEMPA}0 + P\text{TEMPA}1 * y + P\text{TEMPA}2 * y^2$$

$$x = \text{pressure output} - PTCA0 - PTCA1 * t - PTCA2 * t^2$$

$$n = x * PTCB0 / (PTCB0 + PTCB1 * t + PTCB2 * t^2)$$

$$\text{pressurc (psia)} = PA0 + PA1 * n + PA2 * n^2$$

Date, Delta F %FS

21-Nov-03 0.00

