



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA
Website: <http://www.seabird.com>

Phone: (425) 643-9866
FAX: (425) 643-9954
Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 3/15/2007 Description SBE-5T Submersible Pump

Job Number: 46351 Customer Name COMM TECH

SBE Sensor Information:

Model Number: 5T
Serial Number: 4641

Pressure Sensor Information:

Sensor Type: None
Sensor Serial Number: None
Sensor Rating: 0

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test:

Tested By: nd

