



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA
 Website: <http://www.seabird.com>

Phone: (425) 643-9866
 FAX: (425) 643-9954
 Email: seabird@seabird.com

SBE Pressure Test Certificate

Test Date: 5/5/2008 Description SBE-16Plus SeaCat

Job Number: 50181 Customer Name COMM TECH

SBE Sensor Information:

Model Number: 16P

Serial Number: 6047

Pressure Sensor Information:

Sensor Type: DigiQuartz

Sensor Serial Number: 105120

Sensor Rating: 6000

Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 6000 PSI Held For 15 Minutes

Passed Test:

Tested By: ND

