



Sea-Bird Electronics, Inc.

1808 136th Place NE, Bellevue, Washington 98005 USA  
 Website: <http://www.seabird.com>

Phone: (425) 643-9866  
 FAX: (425) 643-9954  
 Email: [seabird@seabird.com](mailto:seabird@seabird.com)

## SBE Pressure Test Certificate

Test Date: 4/9/2008 Description SBE-5T Submersible Pump

Job Number: 50181 Customer Name COMM TECH

### SBE Sensor Information:

Model Number: 5T  
 Serial Number: 4951

### Pressure Sensor Information:

Sensor Type: None  
 Sensor Serial Number: None  
 Sensor Rating: 0

### Pressure Test Protocol:

Low Pressure Test: 40 PSI Held For 15 Minutes

High Pressure Test: 10000 PSI Held For 15 Minutes

Passed Test:

Tested By: FS

